

**Annex to the accreditation certificate  
N°: 2/004  
Standard: ISO/IEC 17025:2017**

**Organism:**

**NATO Support and Procurement Agency  
Agence OTAN de Soutien et d'Acquisition  
(NSPA)**  
**NATO Reference Laboratory  
Laboratoire de Référence de l'OTAN**

11, rue de la Gare

L-8325 Capellen

---

**Version of the technical appendix: 03  
of the 02 March 2022**

**Date of issue: 18 March 2019**

**Expiry date: 18 March 2024**

**Document validated by:**



Dominique Ferrand  
Head of OLAS department

**ILNAS**



# OFFICE LUXEMBOURGEOIS D'ACCREDITATION ET DE SURVEILLANCE

<b>Laboratory:</b>	NATO Support and Procurement Agency (NSPA) – NATO Reference Laboratory	<b>Standard:</b>	ISO/IEC 17025
<b>Contact:</b>	NSPA Calibration (LD-C)	<b>Accreditation n° :</b>	2/004
<b>Street:</b>	11, rue de la Gare	<b>Version:</b>	03
<b>Town:</b>	L-8325 Capellen		
<b>Country:</b>	G.D. Luxembourg		
<b>Telephone:</b>	+352 3063 6144		
<b>Fax:</b>	+352 307255		
<b>E-mail:</b>	calibration@namsa.nato.int		

## Accreditation scope for a calibration laboratory

General domain: CAL1 – Electricity				
Technical domains: CAL1.1 – Voltage				
Objects, instruments subjected to calibration	Characteristics or measured properties	Calibration methods (ex. published, adapted, internally validated)	Measuring range	Calibration and measurement capability (CMC) Expanded uncertainty (k=2) $E_x$ : measured voltage
CAL1.1.1 – Voltage measurers – Direct current (DCV)	Voltage	CMC-11 Substitution of divided voltages	100 mV	0,25 $\mu$ V
		CMC-12 ratio measurements	1 V	1,8 $\mu$ V
			1,018 V	1,8 $\mu$ V
			10 V	8,0 $\mu$ V
		CMC-11 Substitution of divided voltages + ratio measurements	$0 \text{ mV} \leq E_x < 200 \text{ mV}$	$2,0 \cdot 10^{-6} E_x + 0,25 \mu\text{V}$
		CMC-13 Substitution of divided voltages + ratio measurements	$200 \text{ mV} \leq E_x < 2 \text{ V}$	$1,2 \cdot 10^{-6} E_x + 0,50 \mu\text{V}$
			$2 \text{ V} \leq E_x < 20 \text{ V}$	$1,0 \cdot 10^{-6} E_x + 2,5 \mu\text{V}$
			$20 \text{ V} \leq E_x < 200 \text{ V}$	$1,5 \cdot 10^{-6} E_x + 15 \mu\text{V}$
			$200 \text{ V} \leq E_x \leq 1000 \text{ V}$	$2,0 \cdot 10^{-6} E_x + 250 \mu\text{V}$
CAL1.1.2 – Voltage measurers – Alternative current (ACV)	Voltage	CMC-31 Direct measurement with a voltmeter and an AC/DC transfer standard	Fixed points	10 Hz to 1 MHz See <b>Table 1: Matrix CMC-31-fix</b>
			Variable ranges	10 Hz to 1 MHz See <b>Table 2: Matrix CMC-31-var</b>

Technical domains: CAL1.2 – Amperage				
Objects, instruments subjected to calibration	Characteristics or measured properties	Calibration methods (ex. published, adapted, internally validated)	Measuring range $I_x$ : measured current	Calibration and measurement capability (CMC) Expanded uncertainty (k=2) $I_x$ : measured current
CAL1.2.1 – Current measurers and generators – Direct current (DCI)	Current	CMC-21 Measurement of voltage across a standard shunt in series with UUT	0 $\mu\text{A} < I_x \leq \mu\text{A}$	$8,0 \cdot 10^{-6} I_x + 50 \text{ pA}$
			1 $\mu\text{A} < I_x \leq 10 \mu\text{A}$	$12 \cdot 10^{-6} I_x + 60 \text{ pA}$
			10 $\mu\text{A} < I_x \leq 100 \mu\text{A}$	$8,0 \cdot 10^{-6} I_x + 0,4 \text{ nA}$
			100 $\mu\text{A} < I_x \leq 1 \text{ mA}$	$6,0 \cdot 10^{-6} I_x + 5,0 \text{ nA}$
			1 $\text{mA} < I_x \leq 10 \text{ mA}$	$6,0 \cdot 10^{-6} I_x + 40 \text{ nA}$
			10 $\text{mA} < I_x \leq 100 \text{ mA}$	$6,0 \cdot 10^{-6} I_x + 0,5 \mu\text{A}$
			100 $\text{mA} < I_x \leq 1 \text{ A}$	$12 \cdot 10^{-6} I_x + 20 \mu\text{A}$
		CMC-21 Measurement of voltage across a standard shunt	1 $\text{A} < I_x \leq 10 \text{ A}$	$12 \cdot 10^{-6} I_x + 40 \mu\text{A}$
			10 $\mu\text{A} < I_x \leq 100 \mu\text{A}$	$8,0 \cdot 10^{-6} I_x + 0,3 \text{ nA}$
			100 $\mu\text{A} < I_x \leq 1 \text{ mA}$	$6,0 \cdot 10^{-6} I_x + 15 \text{ nA}$
			1 $\text{mA} < I_x \leq 10 \text{ mA}$	$5,0 \cdot 10^{-6} I_x + 40 \text{ nA}$
			10 $\text{mA} < I_x \leq 100 \text{ mA}$	$6,0 \cdot 10^{-6} I_x + 0,4 \mu\text{A}$
			100 $\text{mA} < I_x \leq 1 \text{ A}$	$10 \cdot 10^{-6} I_x + 5,0 \mu\text{A}$
			1 $\text{A} < I_x \leq 10 \text{ A}$	$10 \cdot 10^{-6} I_x + 50 \mu\text{A}$
CAL1.2.2 – Current measurers – Alternative current (ACI)	Current	CMC-41 Measurement of voltage across a standard shunt with a voltmeter and an AC/DC transfer standard	Fixed points	10 Hz to 10 kHz
			Variable ranges	10 Hz to 10 kHz
			See <b>Table 3: Matrix CMC-41-fix</b>	
			See <b>Table 4: Matrix CMC-41-var</b>	

Technical domains: CAL1.3 – Resistance				
Objects, instruments subjected to calibration	Characteristics or measured properties	Calibration methods (ex. published, adapted, internally validated)	Measuring range $R_x$ : measured resistance	Calibration and measurement capability (CMC) Expanded uncertainty ( $k=2$ ) $R_x$ : measured resistance
CAL1.3.1 – Resistance Range Calibration	Resistance	- CMC-05 - Direct Comparison	$100 \mu\Omega < R_x \leq 1 \text{m}\Omega$	$2,5 \cdot 10^{-6}R_x + 0,3 \mu\Omega$
			$1 \text{m}\Omega < R_x \leq 100 \text{ m}\Omega$	$2,0 \cdot 10^{-6}R_x$
			$100 \text{ m}\Omega < R_x \leq 100 \Omega$	$1,5 \cdot 10^{-6}R_x$
			$100 \Omega < R_x \leq 10 \text{k}\Omega$	$1,5 \cdot 10^{-6}R_x$
			$10 \text{k}\Omega < R_x \leq 100 \text{k}\Omega$	$2,0 \cdot 10^{-6}R_x$
			$100 \text{k}\Omega < R_x \leq 1 \text{M}\Omega$	$7,0 \cdot 10^{-6}R_x$
CAL1.3.3 – Resistance measurers	Resistance	- CMC-01 - Direct Measurement	$1 \text{m}\Omega$	$8,0 \cdot 10^{-6}R_x$
			$10 \text{ m}\Omega$	$6,0 \cdot 10^{-6}R_x$
			$100 \text{ m}\Omega$	$3,0 \cdot 10^{-6}R_x$
			$1 \Omega$	$1,0 \cdot 10^{-6}R_x$
			$10 \Omega$	$1,0 \cdot 10^{-6}R_x$
			$100 \Omega$	$1,5 \cdot 10^{-6}R_x$
			$1 \text{k}\Omega$	$2,0 \cdot 10^{-6}R_x$
			$10 \text{k}\Omega$	$2,0 \cdot 10^{-6}R_x$
			$100 \text{k}\Omega$	$6,0 \cdot 10^{-6}R_x$
			$1 \text{M}\Omega$	$8,0 \cdot 10^{-6}R_x$
			$10 \text{ M}\Omega$	$8,0 \cdot 10^{-6}R_x$
			$100 \text{ M}\Omega$	$30 \cdot 10^{-6}R_x$
			$1 \text{G}\Omega$	$100 \cdot 10^{-6}R_x$

## General domain: CAL2 – Time / Frequency

### Technical domains: CAL2.1 – Frequency Counters

Objects, instruments subjected to calibration	Characteristics or measured properties	Calibration methods (ex. published, adapted, internally validated)	Measuring range $F_x$ : measured frequency	Calibration and measurement capability (CMC) <sup>a</sup>	
				Expanded uncertainty ( $k=2$ )	t <sup>b</sup>
			1 mHz $\leq F_x < 10$ mHz	2,9 · 10 <sup>-12</sup>	5,8 · 10 <sup>-11</sup>
			10 mHz $\leq F_x < 100$ mHz		1,9 · 10 <sup>-11</sup>
			100 mHz $\leq F_x < 1$ Hz	6,8 · 10 <sup>-12</sup>	1,9 · 10 <sup>-10</sup>
			1 Hz $\leq F_x < 10$ Hz	4,2 · 10 <sup>-12</sup>	5,7 · 10 <sup>-9</sup>
CAL 2.1.5 – Multimeters / Calibrators – Square wave	Frequency	Measured using a frequency counter with a Caesium frequency standard <sup>c</sup>	10 Hz $\leq F_x < 100$ Hz	2,3 · 10 <sup>-12</sup>	2,1 · 10 <sup>-11</sup>
			100 Hz $\leq F_x < 1$ kHz		1,2 · 10 <sup>-11</sup>
			1 kHz $\leq F_x < 10$ kHz	3,8 · 10 <sup>-12</sup>	1,9 · 10 <sup>-10</sup>
			10 kHz $\leq F_x < 100$ kHz		5,9 · 10 <sup>-11</sup>
			100 kHz $\leq F_x < 1$ MHz	1,1 · 10 <sup>-11</sup>	2,4 · 10 <sup>-11</sup>
			1 MHz $\leq F_x < 12,4$ GHz		1,7 · 10 <sup>-11</sup>
			12,4 GHz $\leq F_x < 26,5$ GHz	-	1,6 · 10 <sup>-11</sup>
				-	2,1 · 10 <sup>-10</sup>

<sup>a</sup> Expanded uncertainties are given in dimensionless values

<sup>b</sup> Gate time of the measurement expressed in second (s)

<sup>c</sup> The accuracy of the Caesium frequency standard, traceable up to national standards, is  $2 \cdot 10^{12}$  ( $k=2$ ).  
Certificat n° 2/004

Objects, instruments subjected to calibration	Characteristics or measured properties	Calibration methods (ex. published, adapted, internally validated)	Measuring range $F_x$ : measured frequency	Calibration and measurement capability (CMC) <sup>d</sup>		
				Expanded uncertainty (k=2)		
CAL 2.1.6 – Multimeters / Calibrators – Sine wave	Frequency	Measured using a frequency counter with a Caesium frequency standard <sup>f</sup>	1 mHz ≤ $F_x$ < 10 mHz	1000 s	100 s	10 s
			10 mHz ≤ $F_x$ < 100 mHz	1,4 · 10 <sup>-3</sup>	-	-
			100 mHz ≤ $F_x$ < 1 Hz	4,3 · 10 <sup>-5</sup>	1,4 · 10 <sup>-3</sup>	-
			1 Hz ≤ $F_x$ < 10 Hz	1,4 · 10 <sup>-6</sup>	4,3 · 10 <sup>-5</sup>	1,4 · 10 <sup>-3</sup>
			10 Hz ≤ $F_x$ < 100 Hz	4,3 · 10 <sup>-8</sup>	1,4 · 10 <sup>-6</sup>	4,3 · 10 <sup>-5</sup>
			100 Hz ≤ $F_x$ < 1 kHz	1,4 · 10 <sup>-9</sup>	4,3 · 10 <sup>-8</sup>	1,4 · 10 <sup>-6</sup>
			1 kHz ≤ $F_x$ < 10 kHz	4,3 · 10 <sup>-11</sup>	1,4 · 10 <sup>-9</sup>	4,3 · 10 <sup>-8</sup>
			10 kHz ≤ $F_x$ < 100 kHz	2,7 · 10 <sup>-12</sup>	4,3 · 10 <sup>-11</sup>	1,4 · 10 <sup>-9</sup>
			100 kHz ≤ $F_x$ < 1 MHz	2,3 · 10 <sup>-12</sup>	4,0 · 10 <sup>-12</sup>	4,4 · 10 <sup>-11</sup>
			1 MHz ≤ $F_x$ < 12,4 GHz		3,8 · 10 <sup>-12</sup>	1,1 · 10 <sup>-11</sup>
			12,4 GHz ≤ $F_x$ < 26,5 GHz	-	-	4,6 · 10 <sup>-11</sup>

<sup>d</sup> Expanded uncertainties are given in dimensionless values

<sup>e</sup> Gate time of the measurement expressed in second (s)

<sup>f</sup> The accuracy of the Caesium frequency standard, traceable up to national standards, is  $2 \cdot 10^{12}$  (k=2).

**Table 1: Matrix CMC-31-fix**

		Expanded Relative Uncertainty of Measurement U in $10^{-6}$ at Frequencies of:												
Range	Input Voltage	10 Hz	20 Hz	40 Hz	120 Hz	1 kHz	10 kHz	20 kHz	50 kHz	100 kHz	300 kHz	500 kHz	800 kHz	1 MHz
22 mV	2 mV	730	690	630	620	650	610	620	660	610	660	780	750	810
	6 mV	290	230	240	230	240	240	240	250	260	310	410	470	600
	10 mV	220	250	190	190	190	190	190	190	210	280	380	510	610
	20 mV	110	77	77	77	81	74	76	76	88	170	270	410	510
	20 mV	110	76	76	76	72	77	74	74	88	170	270	410	500
220 mV	60 mV	54	38	38	39	39	39	39	39	45	74	160	270	400
	100 mV	35	21	21	23	20	21	22	34	52	150	250	390	520
	200 mV	26	17	16	16	15	15	18	28	51	150	250	390	490
	200 mV	26	17	16	16	14	15	17	28	51	150	250	390	490
	600 mV	22	15	15	15	9	10	14	27	51	150	250	390	490
2.2 V	600 mV	30	15	15	15	9	14	17	28	51	150	250	390	490
	1 V	30	11	6	8	8	14	16	28	51	150	250	390	490
	2 V	33	11	8	8	10	15	16	28	51	150	250	390	490
	2 V	27	12	8	8	8	10	14	27	50	150	250	390	490
	6 V	27	11	8	8	6	10	14	27	51	150	250	390	490
22 V	6 V	28	11	6	8	6	9	12	26	50	150	250	390	490
	10 V	33	14	6	6	6	9	12	26	50	150	250	390	490
	20 V	28	11	9	9	10	11	14	27	50	150	250	390	490
	20 V	26	14	9	9	9	10	14	32	50	150	250	390	490
	60 V	32	14	10	10	11	12	16	32	53	150	-	-	-
220 V	60 V	28	12	11	10	11	11	15	27	52	150	-	-	-
	100 V	28	15	15	15	16	18	22	36	63	-	-	-	-
	200 V	35	17	16	16	17	18	20	35	64	-	-	-	-
	200 V	34	16	16	16	15	16	18	29	62	-	-	-	-
	600 V	-	-	16	16	16	17	18	30	71	-	-	-	-
1000 V	1000 V	-	-	22	17	16	17	18	-	-	-	-	-	-

Table 2: Matrix CM-C-31-var

		Expanded Relative Uncertainty of Measurement U in $10^{-6}$ at Frequencies of:											
Range	Input Voltage	10 -20 Hz	20 -40 Hz	40-120 Hz	120 Hz - 1 kHz	1 - 10 kHz	10 -20 kHz	20 -50 kHz	50 -100 kHz	100 -300 kHz	300 - 500 kHz	500 - 800 kHz	800 kHz - 1 MHz
22 mV	2 - 6 mV	730	690	630	650	650	620	660	660	660	780	780	810
	6 - 10 mV	290	260	250	250	250	250	250	260	310	410	510	610
	10 - 20 mV	260	260	190	190	190	200	210	280	380	510	610	
	20 - 60 mV	110	90	90	90	90	90	80	100	170	280	410	520
220 mV	60 - 100 mV	62	47	48	48	48	48	53	80	170	270	400	520
	100 - 200 mV	46	35	36	36	35	36	45	59	160	250	390	490
	200 - 600 mV	39	34	33	33	33	34	40	58	150	250	390	490
2.2 V	600 mV - 1 V	42	33	33	33	32	34	40	58	150	250	390	490
	1 - 2 V	44	30	29	30	33	33	40	58	150	250	390	490
	7 V	2 - 6 V	39	32	29	29	30	32	39	58	150	250	390
22 V	6 - 10 V	44	32	29	29	30	32	39	58	150	250	390	490
	10 - 20 V	44	32	30	30	30	32	39	58	150	250	390	490
	70 V	20 - 60 V	42	32	30	30	32	33	42	60	160	-	-
220 V	60 - 100 V	40	33	33	33	34	36	46	69	-	-	-	-
	100 - 200 V	46	34	33	34	34	36	46	70	-	-	-	-
	200 - 600 V	-	-	33	33	34	34	42	77	-	-	-	-
1000 V	600 - 1000 V	-	-	36	34	34	-	-	-	-	-	-	-

**Table 3: Matrix CMC-41-fix**

Input Current	Expanded Relative Uncertainty of Measurement I in $10^{-6}$ at Frequencies of:										
	10 Hz	20 Hz	30 Hz	40 Hz	55 Hz	400 Hz	500 Hz	1 kHz	2 kHz	5 kHz	10 kHz
10 mA	69	51	68	49	50	48	50	49	69	51	53
30 mA	73	70	74	70	71	74	74	69	74	71	74
100 mA	59	58	58	54	55	53	59	52	56	57	51
300 mA	110	110	100	100	100	93	89	91	120	92	92
1 A	71	64	64	61	59	64	65	59	68	63	71
3 A	120	130	96	110	90	91	94	95	99	110	92
10 A	74	73	73	69	69	67	77	68	74	76	72

**Table 4: Matrix CMC-41-var**

Input Range	Expanded Relative Uncertainty of Measurement I in $10^{-6}$ at Frequencies of:									
	10 -20 Hz	20 - 30 Hz	30 - 40 Hz	40 - 55 Hz	55 - 400 Hz	400 - 500 Hz	0.5 - 1 kHz	1 - 2 kHz	2 - 5 kHz	5 - 10 kHz
5 mA - 10 mA	73	72	72	56	56	55	55	73	73	58
10 mA - 30 mA	77	78	78	75	77	77	77	78	78	78
30 mA - 100 mA	64	32	32	30	30	32	32	31	31	31
100 mA - 300 mA	110	110	110	110	110	96	94	120	120	95
0.3 A - 1 A	75	68	68	66	69	70	70	73	73	75
1 A - 3 A	140	140	110	110	95	97	98	110	110	110
3 A - 10 A	78	77	77	73	73	81	81	78	80	80